

D1008

Palm Micro Centrifuge



Applications

This centrifuge is ideal for quick spins, microfiltration of samples, cell separations and many other routine laboratory procedures.

Features

- Easy to operate with dual start/stop function – Close lid for a spin speed up to 7000 rpm
- Smooth and quiet operation (noise level $\leq 45\text{dB}$)
- Accelerates and brakes in seconds
- Safe, durable and robust rotor with clamp locking design
- Compatible with 2mLx8 and PCR8x4 rotors
- Quick-change rotor without any tool

Specifications	D1008
Max. Speed	7000rpm
Max. RCF	2680 $\times$ g
Rotor Capacity	0.2/0.5/1.5/2mL $\times$ 8 0.2mL $\times$ 32PCR tubes 0.2mL $\times$ 4PCR8-strips
Run Time	Continuous operation
Driving Motor	DC motor
Power Requirements	AC100-240V/50Hz/60Hz 20W
Noise Level	$\leq 45\text{dB}$
Dimension [D $\times$ W $\times$ H]	160x170x122mm
Weight	0.5kg

Available rotors

A4-PCR8P Cat.No.19201165

Max. speed:7000rpm
Max. RCF:2680 $\times$ g
Rotor capacity:
0.2mL $\times$ 32PCR tubes,
0.2mL $\times$ 4PCR8-strips
Rotor material: High strength plastic



A8-2P Cat.No.19201164

Max. speed:7000rpm
Max. RCF:2680 $\times$ g
Rotor capacity:
2mL/1.5mL $\times$ 8, 0.2mL $\times$ 8, 0.5mL $\times$ 8
Rotor material: High strength plastic
Available rotor adapters: SA02P2, SA05P2



Available adapters

SA02P2 Cat.No.19400029

0.2mL rotor adapter
Used with A8-2P rotor
8pcs/pk



SA05P2 Cat.No.19400026

0.5mL rotor adapter
Used with A8-2P rotor
8pcs/pk



Quiet operation



Vibration free



WolfLabs

Pricing on any accessories shown can be found by keying the part number into the search box on our website.

The specifications listed in this brochure are subject to change by the manufacturer and therefore cannot be guaranteed to be correct. If there are aspects of the specification that must be guaranteed, please provide these to our sales team so that details can be confirmed.

www.wolflabs.co.uk

Tel : 01759 301142

Fax : 01759 301143

sales@wolflabs.co.uk

Please contact us if this literature doesn't answer all your questions.